

CARDAN LABORATORIES, INC.  
PO BOX 541  
634 14TH STREET  
NORTHERN CAMBRIA, PA 15714-0541  
PH: 814-948-8650 FAX: 814-948-5893  
Coal, Water, Gas, Rock, Dust, Mine Air, Etc.  
DEP CERTIFICATION 11-139

09/30/2016

PATTON BORO  
P.O. BOX 175  
PATTON, PA 16668

LABORATORY NUMBER: 5307

SOURCE: ID 703

SAMPLED BY: KRT DATE SAMPLED: 09/28/2016 09:00  
DATE REC'D...: 09/28/2016 11:55

B A C T E R I O L O G I C A L E X A M I N A T I O N

DATE STARTED: 09/28/2016 13:45 ANALYST: SS  
DATE ENDED...: 09/29/2016 13:40 ANALYST: SS

TOTAL COLIFORM - MEMBRANE FILTRATION

TOTAL COLIFORM (M-ENDO AT 35.0 DEGREES C. FOR 24 HRS)

VOLUME OF SAMPLE: 100 ML

TOTAL COLIFORM: ABSENCE  
METHOD: SM9222B

THIS TEST HAS INDICATED THAT THIS WATER IS ACCEPTABLE.

No problems were encountered during analysis of this workorder,  
except if noted in this report.

CARDAN LABORATORIES, INC.

BY: *Scott Semelange*